

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 21 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130122</u> ✓
Name		<u>Tim Tuckwood</u>	ID# <u>3446</u> Date <u>8-18-09</u> ✓
A	Agency	<u>Alaska State Troopers</u>	Phone # <u>895-4800</u>
Instrument Location		<u>Delta Courthouse</u> ✓	
Alco S/N		<u>X301596</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>500</u>
B	Alco Test Value Average	<u>.081</u> ✓ 1 st Alco	<u>.081</u> ✓ 2 nd Alco
Signature		<u>Tim Tuckwood</u> ✓	<u>OB</u> <u>8/24/09</u>
(OVER)			


I, Nita J. Bolz, after being first duly sworn, depose and state as follows (State as follows in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 29th day of Sept., 2009.

Carolyn M. Noland
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130122 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122 ✓

AUGUST 18, 2009

OPERATOR'S NAME:

TUCKWOOD TIM

OPERATOR'S NUMBER: 3446

SUBJECT'S LAST NAME:

VERIF

SUBJECT'S FIRST NAME/MI :
OF CAL

O.L. #: 12345

DEPT/AGENCY: DJT0

CASE/REPORT: 09067

TEST TYPE: U

ALCO TARGET VALUE: .081 ✓

ALCO S/N: X301596 ✓

— BREATH ANALYSIS —

.081 ADJUSTED FOR 28.76 in
ALCO TARGET .077 08:53
BLANK TEST .000 08:54
INTERNAL STANDARD VERIFIED 08:54
ALCO TV 28.76 in .081 ✓ 08:55
BLANK TEST .000 08:55
SUBJECT SAMPLE .000 08:56
BLANK TEST .000 08:56
ALCO TV 28.76 in .081 ✓ 08:57
BLANK TEST .000 08:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

AUGUST 18, 2009 ✓

TIME 08:51

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 40c
BAROMETER: 28.76 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~
print test